

THIS DATA SHEET IS FOR IDENTIFICATION PURPOSES ONLY. NOMINAL DIMENSIONS IN MM.

TECHNICAL DATA:	ACCESSORIES:	INSTALLATION:
TERMINATION:- NOMINAL OPERATING VOLTAGE 100KV DC - USING AN ESSEX X-RAY APPROVED RECEPTACLE CABLE LENGTH TOLERANCE +/- 100mm CABLE:- VOLTAGE WITHSTAND 100KV DC 30KV (RMS) AC	CLAMPING SLEEVES:— STANDARD FLANGE T20131/A QUICK LOCK T20096/B ANDREX TYPE FLANGE T20045/C INSTALLATION GREASE P00075	WARNING: CABLE SHOULD ONLY BE FITTED BY A TRAINED PERSON. OIL MUST NOT BE USED ON THE RUBBER TERMINATION, ONLY SUITABLE HIGH VOLTAGE GREASE. ENSURE BY ADJUSTING THE CLAMPING SLEEVE THAT WHEN THE TERMINATION IS FULLY INSERTED IN THE SOCKET THERE IS A GAP OF 6mm BETWEEN THE MATING FACES OF THE CLAMPING SLEEVE AND SOCKET BEFORE TIGHTENING THE CLAMPING SLEEVE.
FOR FULL CABLE DATA SEE ESSEX CABLE DATA SHEET 🛆		

ESSEX

Dandong Light Electric Co.,Ltd Rm.1606 Jiadi Square,No.64,Binjiang Middle Rd, Zhenxing, Dandong, Liaoning China T:0415-6123232 F:0415-6123533

		REVISION:	IONS DUE TO CONTINUING PRODUCT DEVELOPMENT ESSEX X-RAY RESERVES THE RIGHT TO M				RESERVES THE RIGHT TO MODIFY DATA WIT	THOUT NOTICE					
1	١٥.	DATE	BY	DRAWN	APPROVED	DATE	ISSUE	SCALE		O NOT TO BE REPRODUCED IN WHOLE OR PART W	IN WHOLE OR PART WITHOUT		
	4	02.01.01	DCN 1405	JDF		12.09.95	7		N.T.S.	PRIOR PERMISSION OF ESSEX X-RAY & MEDICA			
	5	11.05.01	DCN 1455	TITLE	TITLE 160KV EURO STRAIGHT (R24 TYPE)								
	6	11.09.03	DCN 1912		IOUN) DC	1003						
	7	06.03.09	DCN 2905		-)	1005							