

THIS DATA SHEET IS FOR IDENTIFICATION PURPOSES ONLY. NOMINAL DIMENSIONS IN MM.

TECHNICAL DATA:	ACCESSORIES:	INSTALLATION:
TERMINATION:— NOMINAL OPERATING VOLTAGE 210KV DC - USING AN ESSEX X—RAY APPROVED RECEPTACLE CABLE LENGTH TOLERANCE +/- 100mm CABLE:— VOLTAGE WITHSTAND 160KV DC	CLAMPING SLEEVES:- STANDARD FLANGE T20172/A QUICK LOCK T20173	WARNING: CABLE SHOULD ONLY BE FITTED BY A TRAINED PERSON. OIL MUST NOT BE USED ON THE RUBBER TERMINATION, ONLY SUITABLE HIGH VOLTAGE GREASE. ENSURE BY ADJUSTING THE CLAMPING SLEEVE THAT WHEN THE TERMINATION IS FULLY INSERTED IN THE SOCKET THERE IS A GAP OF 6mm BETWEEN THE MATING FACES OF THE CLAMPING SLEEVE
50KV (RMS) AC 点	INSTALLATION GREASE P00075	AND SOCKET BEFORE TIGHTENING THE CLAMPING SLEEVE.

ESSEX

Dandong Light Electric Co.,Ltd Rm.1606 Jiadi Square,No.64,Binjiang Middle Rd, Zhenxing, Dandong, Liaoning China T:0415-6123232 F:0415-6123533

	REVISION	S	DUE TO C	ONTINUING	PRODUCT [DEVELOPMENT	ESSEX	X-RAY	RESERVES	THE RI	IGHT TO	MODIFY	DATA WITHOUT NOTICE
NO.	DATE	BY	DRAWN	APPROVED	DATE	ISSUE	SCALE		О мот то	BE REPR	RODUCED	IN WHOLE	OR PART WITHOUT
4	02.01.01	DCN 1405	JDF		15.09.95	5 7		N.T.S.					Y & MEDICAL EQUIPMENT LTD.
5	11.05.01	DCN 1455	TITLE ,	210KV	SFIFFR	T STR/	AIGHT	(R	30 TYF	oF)			DRAWING No.
6	16.07.01	DCN 1534		21011				(50 111	レ)			DC1005
7	06.03.09	DCN 2905			(CAB	LE TYPE	. 020	142)					D31003