

THIS DATA SHEET IS FOR IDENTIFICATION PURPOSES ONLY. NOMINAL DIMENSIONS IN MM.

TECHNICAL DATA.

IMPORTANT: 3 HOLES M5x10 IN S61006/C & S61008/E ONLY

INISTALLATION!

TECHNICAL DATA.	ACCESSORIES.	INSTALLATION.			
OPERATING VOLTAGE 210KV DC  1. FLANGE MATERIAL ALUMINIUM. TERMINAL MATERIAL BRASS  2. CONTACT CAVITY AND TAPER MUST NOT BE FILLED WITH ANY PETROLEUM BASED OIL.  3. MATERIAL OF INSULATION IS CAPABLE OF CONTINUOUS IMMERSION IN SHELL DIALA B.G. OIL.  ALL TYPES EPOXY RESIN COLOUR GREY \$61006/B STANDARD CONTACTS \$61008/E WIDE BAND CONTACTS	CABLE ASSEMBLY USED WITH THIS RECEPTACLE TO HAVE 210KV SEIFERT TERMINATION.  CLAMPING SLEEVES:—  FLANGE TO BE USED T20172/A  QUICK LOCK T20173	WARNING: THIS SOCKET MUST ONLY BE FITTED BY A TRAINED PERSON. SEE CABLE TERMINATION DATA SHEET DS1005 OR DS1006 FOR FITTING OF CABLE TERMINATION TO RECEPTACLE.			

ACCESSORIES.

ESSEX

Dandong Light Electric Co.,Ltd Rm.1606 Jiadi Square,No.64,Binjiang Middle Rd, Zhenxing, Dandong, Liaoning China T:0415-6123232 F:0415-6123533

	REVISION	S	DUE TO C	ONTINUING	PRODUCT D	EVELOPMEN	T ESSEX	X-RAY	RESERVES	THE RIGH	т то мо	ODIFY	DATA WITHOUT NOTICE
NO.	DATE	BY	DRAWN	APPROVED	DATE	ISSUE	SCALE		О мот то	BE REPROD	UCED IN W	WHOLE	OR PART WITHOUT
5	01.03.04	DCN 2000	JDF		15.07.96	8		N.T.S.	PRIOR F	PERMISSION (	OF ESSEX	X-RAY	& MEDICAL EQUIPMENT LTD.
6	27.02.06	DCN 2432	TITLE 21	NKV SE	IFERT	(R30)	GENE	RATO	R RECI	FPTACI	F		DRAWING No.
7	06.11.07	DCN 2708										/-	DC1011
8	12.02.08	DCN 2746	E551	LX PAR	I NO.	S6100	o/B,	2010	106/6	& 50	1008,	/ <u> </u>	031014