



THIS DATA SHEET IS FOR IDENTIFICATION
PURPOSES ONLY. NOMINAL DIMENSIONS IN MM.

TECHNICAL DATA:

ACCESSORIES:

INSTALLATION:

OPERATING VOLTAGE 75KV DC

1. TERMINAL MATERIAL BRASS

2. MATERIAL OF INSULATION IS CAPABLE OF CONTINUOUS
IMMERSION IN SHELL DIALA B.G. OIL OR EQUIVALENT.

CABLE ASSEMBLY USED WITH THIS RECEPTACLE TO HAVE
A STANDARD 75KV TERMINATION 3 PIN TYPE

FLANGE CAN BE SUPPLIED:—

WITHOUT LOCATION HOLES T20596
WITH 4 LOCATION HOLES T20596/A
WITH 6 LOCATION HOLES T20596/B
WITH KEYLOCK RING T20768

WARNING:
THIS SOCKET MUST ONLY BE FITTED BY A TRAINED
PERSON SEE CABLE TERMINATION DATA SHEET
DS 1020 FOR FITTING OF CABLE TERMINATION.

ESSEX

Dandong Light Electric Co.,Ltd
Rm.1606 Jiadi Square,No.64,Binjiang Middle Rd,
Zhenxing, Dandong, Liaoning China
T:0415-6123232
F:0415-6123533

REVISIONS			DUE TO CONTINUING PRODUCT DEVELOPMENT ESSEX X-RAY RESERVES THE RIGHT TO MODIFY DATA WITHOUT NOTICE					
NO.	DATE	BY	DRAWN	APPROVED	DATE	ISSUE	SCALE	© NOT TO BE REPRODUCED IN WHOLE OR PART WITHOUT PRIOR PERMISSION OF ESSEX X-RAY & MEDICAL EQUIPMENT LTD.
3	15.07.99	DCN 1049	JDF		09.09.96	6	N.T.S.	
4	11.05.01	DCN 1455	TITLE 3 PIN 75KV RECEPTACLE WITH FLANGE					DRAWING No. DS1022
5	29.11.06	DCN 2579						
6	16.06.08	DCN 2790						