



NOTE:

1. TO BE MANUFACTURED TO PROVIDE THE MAXIMUM FLEXIBILITY OF THE FINAL CABLE.
2. ISOLATION BETWEEN SCREENS TO BE NOT LESS THAN 100 MOhms.

ESSEX

Dandong Light Electric Co.,Ltd
Rm.1606 Jiadi Square,No.64,Binjiang Middle Rd,
Zhenxing, Dandong, Liaoning China
T:0415-6123232
F:0415-6123533

REVISIONS			DUE TO CONTINUING PRODUCT DEVELOPMENT ESSEX X-RAY RESERVES THE RIGHT TO MODIFY DATA WITHOUT NOTICE						
NO.	DATE	BY	DRAWN	APPROVED	DATE	ISSUE	SCALE	© NOT TO BE REPRODUCED IN WHOLE OR PART WITHOUT PRIOR PERMISSION OF ESSEX X-RAY & MEDICAL EQUIPMENT LTD.	
2	15.07.99	DCN 1049	TSC		12.6.97	3	N.T.S.		
3	27.01.09	DCN 2895	TITLE SPECIAL 3 CORE CABLE ESSEX PART NO. C9901						DRAWING No. DS1026