

TECHNICAL DATA:	ACCESSORIES:	INSTALLATION:
OPERATING VOLTAGE 160KV DC 1. FLANGE MATERIAL S51007, S51007/A & S51028/A BRASS BRIGHT NICKEL PLATED. S51007/D, S51007/E, S51038 ALUMINIUM. TERMINAL MATERIAL ALL TYPES BRASS.	CABLE ASSEMBLY USED WITH THIS SOCKET TO HAVE TERMINATION 160 KV EURO (R24 TYPE). FLANGE TO BE USED T20131/A	WARNING: THIS SOCKET MUST ONLY BE FITTED BY A TRAINED PERSON SEE CABLE TERMINATION DATA SHEET DS 1002 FOR FITTING OF CABLE TERMINATION
CONTACT CAVITY & TAPER MUST NOT BE FILLED WITH ANY PETROLIUM BASED OIL. MATERIAL OF INSULATION IS CAPABLE OF CONTINUOUS IMMERSION IN SHELL DIALA B.G. OIL OR EQUIVALENT. EPOXY RESIN COLOUR ALL TYPES GREY	S51007 WIDE BAND CONTACTS S51007/A WIDE BAND CONTACTS & TAPPED HOLES S51007/D WIDE BAND CONTACTS S51007/E WIDE BAND CONTACTS & TAPPED HOLES S51028/A WITH TAPPED HOLES S51038 WITH TAPPED HOLES	TO SOCKET.

ESSEX

Dandong Light Electric Co.,Ltd Rm.1606 Jiadi Square,No.64,Binjiang Middle Rd, Zhenxing, Dandong, Liaoning China T:0415-6123232 F:0415-6123533

REVISIONS			DUE TO C	ONTINUING P	RODUCT D	EVELOPMENT	ESSEX	X-RAY	RESERVES	THE R	IGHT TO	MODIFY	DATA WITHOUT NOTICE
NO.	DATE	BY	DRAWN	APPROVED	DATE	ISSUE	SCALE		С мот то	BE REP	RODUCED	IN WHOLE	OR PART WITHOUT
4	01.03.04	DCN 2000	MGM		07.07.98	7		N.T.S.	PRIOR	PERMISSIO	ON OF ES	SSEX X-RA	Y & MEDICAL EQUIPMENT LTD.
5	27.02.06	DCN 2432	TITLE	160KV (R24) EURO GENERATOR RECEPTACLES								DRAWING No.	
6	14.02.07	DCN 2616		,							DS1030		
7	21.02.08	2752	ESSEX PART NO. AS DETAILED							D31030			