



THIS DATA SHEET IS FOR IDENTIFICATION
PURPOSES ONLY. NOMINAL DIMENSIONS IN MM.

TECHNICAL DATA:

OPERATING VOLTAGE 100KV DC

1. TERMINAL MATERIAL BRASS.
2. CONTACT CAVITY & TAPER MUST NOT BE FILLED WITH ANY PETROLIUM BASED OIL.
3. MATERIAL OF INSULATION IS CAPABLE OF CONTINUOUS IMMERSION IN SHELL DIALA B.G. OIL OR EQUIVALENT.

ACCESSORIES:

CABLE ASSEMBLY USED WITH THIS SOCKET TO HAVE
TERMINATION 100 KV EURO (R10 TYPE).

INSTALLATION:

WARNING:
THIS SOCKET MUST ONLY BE FITTED BY A TRAINED
PERSON SEE CABLE TERMINATION DATA SHEET
DS 1007 FOR FITTING OF CABLE TERMINATION
TO SOCKET.

ESSEX

Dandong Light Electric Co.,Ltd
Rm.1606 Jiadi Square, No.64, Binjiang Middle Rd,
Zhenxing, Dandong, Liaoning China
T:0415-6123232
F:0415-6123533

REVISIONS

NO.	DATE	BY
2	14.07.99	DCN 1049
3	11.05.01	DCN 1455
4	08.07.04	DCN 2047

DUE TO CONTINUING PRODUCT DEVELOPMENT ESSEX X-RAY RESERVES THE RIGHT TO MODIFY DATA WITHOUT NOTICE

DRAWN	APPROVED	DATE	ISSUE	SCALE
MGM		8.7.98	4	1:1

© NOT TO BE REPRODUCED IN WHOLE OR PART WITHOUT
PRIOR PERMISSION OF ESSEX X-RAY & MEDICAL EQUIPMENT LTD.

TITLE
100KV (R10) TUBE RECEPTACLE
ESSEX PART NO. S40012

DRAWING No.

DS1031