



THIS DATA SHEET IS FOR IDENTIFICATION
PURPOSES ONLY. NOMINAL DIMENSIONS IN MM.

TECHNICAL DATA:

OPERATING VOLTAGE 160KV DC

1. TERMINAL MATERIAL BRASS.
2. CONTACT CAVITY & TAPER MUST NOT BE FILLED WITH ANY PETROLIUM BASED OIL.
3. MATERIAL OF INSULATION IS CAPABLE OF CONTINUOUS IMMERSION IN SHELL DIALA B.G. OIL OR EQUIVALENT.

ACCESSORIES:

CABLE ASSEMBLY USED WITH THIS SOCKET TO HAVE
TERMINATION 160 KV EURO (R24 TYPE).

INSTALLATION:

WARNING:
THIS SOCKET MUST ONLY BE FITTED BY A TRAINED
PERSON SEE CABLE TERMINATION DATA SHEET
DS 1002 FOR FITTING OF CABLE TERMINATION
TO SOCKET.

ESSEX

Dandong Light Electric Co.,Ltd
Rm.1606 Jiadi Square,No.64,Binjiang Middle Rd,
Zhenxing, Dandong, Liaoning China
T:0415-6123232
F:0415-6123533

REVISIONS

NO.	DATE	BY
2	14.07.99	DCN 1049
3	11.05.01	DCN 1455
4	08.07.04	DCN 2047

DUE TO CONTINUING PRODUCT DEVELOPMENT ESSEX X-RAY RESERVES THE RIGHT TO MODIFY DATA WITHOUT NOTICE

DRAWN	APPROVED	DATE	ISSUE	SCALE
MGM		8.7.98	4	N.T.S.

© NOT TO BE REPRODUCED IN WHOLE OR PART WITHOUT
PRIOR PERMISSION OF ESSEX X-RAY & MEDICAL EQUIPMENT LTD.

TITLE
160KV (R24) TUBE RECEPTACLE
ESSEX PART NO.S50025

DRAWING No.

DS1032