

THIS DATA SHEET IS FOR IDENTIFICATION PURPOSES ONLY. NOMINAL DIMENSIONS IN MM.

TECHNICAL DATA:	ACCESSORIES:	INSTALLATION:
TERMINATION:—  NOMINAL OPERATING VOLTAGE 75KV DC — USING AN ESSEX X—RAY APPROVED RECEPTACLE  CABLE LENGTH TOLERANCE +/— 100mm  SOLID RESIN TERMINATION.  CABLE:— VOLTAGE WITHSTAND 75KV DC 25KV (RMS) AC 6  TERMINATION RESIN FIRE RESISTANT TO UL VO—94  FOR FULL CABLE DATA SEE ESSEX CABLE DATA SHEET 6	PIN REPLACEMENT KIT P00432 INSTALLATION GREASE P00075 SPLIT WASHER T20226 LOCKING RING T20443	WARNING: CABLE SHOULD ONLY BE FITTED BY A TRAINED PERSON. CABLE CAN RETAIN ELECTRICAL CHARGE. DISCHARGE TO EARTH BEFORE HANDLING. A SUITABLE HIGH VOLTAGE GREASE SHOULD BE APPLIED TO THE TERMINATION. ENSURE THAT THE TERMINATION IS FULLY INSERTED IN THE SOCKET BEFORE SECURING THE LOCKING RING. AFTER INSTALLATION FULLY TIGHTEN GRUB SCREWS TO PREVENT UNAUTHORISED DISCONNECTION.

## ESSEX

Dandong Light Electric Co.,Ltd Rm.1606 Jiadi Square,No.64,Binjiang Middle Rd, Zhenxing, Dandong, Liaoning China T:0415-6123232 F:0415-6123533

	REVISION	S	DUE TO C	ONTINUING F	RODUCT DE	EVELOPMENT	ESSEX	X-RAY	RESERVES THE RIGHT TO	MODIFY	DATA WITHOUT NOTICE
NO.	DATE	BY	DRAWN	APPROVED	DATE	ISSUE	SCALE		© NOT TO BE REPRODUCED	IN WHOLE	OR PART WITHOUT
3	05.01.01	DCN 1408	TSC		24.06.99	6		N.T.S.			' & MEDICAL EQUIPMENT LTD.
4	11.05.01	DCN 1455	3 PIN 75KV TERMINATION ESSEX 2000 (CABLE TYPE C2214)								DRAWING No.
5	29.11.06	DCN 2583									DC1037
6	09.03.09	DCN 2905			(CADI		_ \( \( \Z \).	214)			D31037