



### TECHNICAL DATA:

#### TERMINATION:-

NOMINAL OPERATING VOLTAGE 210KV DC  
 – USING AN ESSEX X-RAY APPROVED RECEPTACLE  
 CABLE LENGTH TOLERANCE +/- 100mm

CABLE:- VOLTAGE WITHSTAND 250KV DC  
 75KV (RMS) AC 

FOR FULL CABLE DATA SEE ESSEX CABLE DATA SHEET 

### ACCESSORIES:

#### CLAMPING SLEEVES:-

STANDARD FLANGE T20172/A  
 QUICK LOCK T20173

INSTALLATION GREASE P00075

### INSTALLATION:

#### WARNING:

CABLE SHOULD ONLY BE FITTED BY A TRAINED PERSON.  
 OIL MUST NOT BE USED ON THE RUBBER TERMINATION, ONLY SUITABLE HIGH VOLTAGE GREASE.  
 ENSURE BY ADJUSTING THE CLAMPING SLEEVE THAT WHEN THE TERMINATION IS FULLY INSERTED IN THE SOCKET THERE IS A GAP OF 6mm BETWEEN THE MATING FACES OF THE CLAMPING SLEEVE AND SOCKET BEFORE TIGHTENING THE CLAMPING SLEEVE.

# ESSEX

Dandong Light Electric Co.,Ltd  
 Rm.1606 Jiadi Square, No.64, Binjiang Middle Rd,  
 Zhenxing, Dandong, Liaoning China  
 T:0415-6123232  
 F:0415-6123533

### REVISIONS

NO.	DATE	BY
2	16.07.01	DCN 1534
3	09.03.09	DCN 2905

DUE TO CONTINUING PRODUCT DEVELOPMENT ESSEX X-RAY RESERVES THE RIGHT TO MODIFY DATA WITHOUT NOTICE

DRAWN	APPROVED	DATE	ISSUE	SCALE
TC		07.03.01	3	N.T.S.

© NOT TO BE REPRODUCED IN WHOLE OR PART WITHOUT PRIOR PERMISSION OF ESSEX X-RAY & MEDICAL EQUIPMENT LTD.

### TITLE

210 SEIFERT STRAIGHT (R30 TYPE)  
 (CABLE TYPE C2236)

### DRAWING No.

DS1055