

THIS DATA SHEET IS FOR IDENTIFICATION PURPOSES ONLY. NOMINAL DIMENSIONS IN MM.

TECHNICAL DATA.

TECHNICAL DATA,	ACCESSONIES.	INSTALLATION.
1. FLANGE MATERIAL S/1001, S/1001/F BRASS BRIGHT NICKEL PLATED. S71001/A, S71001/AF ALUMINIUM TERMINAL MATERIAL ALL TYPES BRASS 2. CONTACT CAVITY AND TAPER MUST NOT BE FILLED WITH ANY PETROLEUM BASED OIL. 3. MATERIAL OF INSULATION IS CAPABLE OF CONTINUOUS IMMERSION IN SHELL DIALA B.G. OIL OR EQUIVALENT.	CABLE ASSEMBLY USED WITH THIS SOCKET TO HAVE TERMINATION 225KV (R28 TYPE)  SUITABLE FLANGES: T20405 T20289  S71001/F WITH SHORTING PLATE S71001/AF WITH SHORTING PLATE	WARNING: THIS SOCKET MUST ONLY BE FITTED BY A TRAINED PERSON. SEE CABLE TERMINATION DATA SHEET DS1041 FOR FITTING OF CABLE TERMINATION TO SOCKET.

ACCESSORIES.

## ESSEX

Dandong Light Electric Co.,Ltd Rm.1606 Jiadi Square, No.64, Binjiang Middle Rd, Zhenxing, Dandong, Liaoning China T:0415-6123232 F:0415-6123533

		REVISIONS	S	DUE TO C	ONTINUING	PRODUCT D	EVELOPMENT	ESSEX	X-RAY	RESERVES THE RIGHT TO	MODIFY	DATA WITHOUT NOTICE	
	NO.	DATE	BY	DRAWN	APPROVED	DATE	ISSUE	SCALE		O NOT TO BE REPRODUCED	IN WHOLE	OR PART WITHOUT	1
	2	14.02.07	DCN 2616	MGM		16.11.05	2		N.T.S.			/ & MEDICAL EQUIPMENT LTD.	
[				TITLE	TILE 225KV (R28) GENRATOR RECEPTACLE						DRAWING No.	1	

ESSEX PART NO. S71001

DS1070

INICTALLATIONI.