



THIS DATA SHEET IS FOR IDENTIFICATION PURPOSES ONLY. NOMINAL DIMENSIONS IN MM.

TECHNICAL DATA:

OPERATING VOLTAGE 210KV DC

1. TERMINAL MATERIAL BRASS

2. CONTACT CAVITY & TAPER MUST NOT BE FILLED WITH ANY PETROLIUM BASED OIL.

3. MATERIAL OF INSULATION IS CAPABLE OF CONTINUOUS IMMERSION IN SHELL DIALA B.G. OIL OR EQUIVALENT.

ACCESSORIES:

CABLE ASSEMBLY USED WITH THIS RECEPTACLE TO HAVE 210KV R30 TERMINATION.

INSTALLATION:

WARNING:
THIS SOCKET MUST ONLY BE FITTED BY A TRAINED PERSON SEE CABLE TERMINATION DATA SHEET DS 1005 OR DS1006 FOR FITTING OF CABLE TERMINATION TO RECEPTACLE.

ESSEX

Dandong Light Electric Co.,Ltd
Rm.1606 Jiadi Square,No.64,Binjiang Middle Rd,
Zhenxing, Dandong, Liaoning China
T:0415-6123232
F:0415-6123533

REVISIONS

NO.	DATE	BY

DUE TO CONTINUING PRODUCT DEVELOPMENT ESSEX X-RAY RESERVES THE RIGHT TO MODIFY DATA WITHOUT NOTICE

DRAWN	APPROVED	DATE	ISSUE	SCALE
MGM		06.11.09	1	N.T.S.

© NOT TO BE REPRODUCED IN WHOLE OR PART WITHOUT PRIOR PERMISSION OF ESSEX X-RAY & MEDICAL EQUIPMENT LTD.

TITLE
210KV (R30) 4 PIN TUBE RECEPTACLE
ESSEX PART NO. S61001/A

DRAWING No.

DS1077